



Lancaster Laboratories
Environmental

2425 New Holland Pike, Lancaster, PA 17601 • 717-656-2300 • Fax: 717-656-6766 • www.EurofinsUS.com/LancLabsEnv

Analysis Report

Sample Description:

Ex. 6 - Personal Privacy
Wolverine World Wide

Tetra Tech, Inc.
ELLE Sample #:
ELLE Group #:
Matrix: Water

Ex. 6 - Personal Privacy

Project Name:

Wolverine World Wide Tannery

Submission Date/Time:

12/29/2017 10:00

Collection Date/Time:

12/27/2017 17:19

SDG#:

Ex. 6 - Personal Privacy

CAT No.	Analysis Name	CAS Number	Result	Method Detection Limit	Dilution Factor
Misc. Organics		EPA 537 Version 1.1	ng/l	ng/l	
14070	NEtFOSAA	2991-50-6	N.D.	2	1
	NEtFOSAA is the acronym for N-ethyl perfluorooctanesulfonamidoacetic Acid.				
14070	NMeFOSAA	2355-31-9	N.D.	2	1
	NMeFOSAA is the acronym for N-methyl perfluorooctanesulfonamidoacetic Acid.				
14070	Perfluorobutanesulfonate	375-73-5	N.D.	2	1
14070	Perfluorodecanoic acid	335-76-2	N.D.	2	1
14070	Perfluorododecanoic acid	307-55-1	N.D.	2	1
14070	Perfluoroheptanoic acid	375-85-9	N.D.	2	1
14070	Perfluorohexanesulfonate	355-46-4	N.D.	2	1
14070	Perfluorohexanoic acid	307-24-4	N.D.	2	1
14070	Perfluorononanoic acid	375-95-1	N.D.	2	1
14070	Perfluoro-octanesulfonate	1763-23-1	N.D.	2	1
14070	Perfluorooctanoic acid	335-67-1	N.D.	2	1
14070	Perfluorotetradecanoic acid	376-06-7	N.D.	3	1
14070	Perfluorotridecanoic acid	72629-94-8	N.D.	2	1
14070	Perfluoroundecanoic acid	2058-94-8	N.D.	2	1

Sample Comments

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
14070	Full List PFAS - DW	EPA 537 Version 1.1	1	18005007	01/16/2018 00:17	Marissa C Drexinger	1
14381	DW PFAS Prep	EPA 537 Version 1.1	1	18005007	01/05/2018 07:35	Pamela Rothharpt	1



Lancaster Laboratories
Environmental

Analysis Report

2425 New Holland Pike, Lancaster, PA 17601 • 717-656-2300 • Fax: 717-656-6766 • www.EurofinsUS.com/LancLabsEnv

Sample Description: Ex. 6 - Personal Privacy
Wolverine World Wide

Tetra Tech, Inc.
ELLE Sample #:
ELLE Group #: Ex. 6 - Personal Privacy
Matrix: Water

Project Name: Wolverine World Wide Tannery

Submittal Date/Time: 12/29/2017 10:00

Collection Date/Time: 12/27/2017 17:23

SDG#: Ex. 6 - Personal Privacy

CAT No.	Analysis Name	CAS Number	Result	Method Detection Limit	Dilution Factor
Misc. Organics		EPA 537 Version 1.1	ng/l	ng/l	
14070	NEtFOSAA	2991-50-6	N.D.	2	1
	NEtFOSAA is the acronym for N-ethyl perfluorooctanesulfonamidoacetic Acid.				
14070	NMeFOSAA	2355-31-9	N.D.	2	1
	NMeFOSAA is the acronym for N-methyl perfluorooctanesulfonamidoacetic Acid.				
14070	Perfluorobutanesulfonate	375-73-5	N.D.	2	1
14070	Perfluorodecanoic acid	335-76-2	N.D.	2	1
14070	Perfluorododecanoic acid	307-55-1	N.D.	2	1
14070	Perfluoroheptanoic acid	375-85-9	N.D.	2	1
14070	Perfluorohexanesulfonate	355-46-4	N.D.	2	1
14070	Perfluorohexanoic acid	307-24-4	N.D.	2	1
14070	Perfluorononanoic acid	375-95-1	N.D.	2	1
14070	Perfluoro-octanesulfonate	1763-23-1	20	2	1
14070	Perfluorooctanoic acid	335-67-1	N.D.	2	1
14070	Perfluorotetradecanoic acid	376-06-7	N.D.	3	1
14070	Perfluorotridecanoic acid	72629-94-8	N.D.	2	1
14070	Perfluoroundecanoic acid	2058-94-8	N.D.	2	1

Sample Comments

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
14070	Full List PFAS - DW	EPA 537 Version 1.1	1	18010003	01/12/2018 21:54	Marissa C Drexinger	1
14381	DW PFAS Prep	EPA 537 Version 1.1	2	18010003	01/10/2018 09:00	Pamela Rothharpt	1



Lancaster Laboratories
Environmental

Analysis Report

2425 New Holland Pike, Lancaster, PA 17601 • 717-656-2300 • Fax: 717-656-6766 • www.EurofinsUS.com/LancLabsEnv

Sample Description:

Ex. 6 - Personal Privacy

~~Wolverine World Wide~~

Tetra Tech, Inc.
ELLE Sample #:
ELLE Group #:
Matrix: Water

Ex. 6 - Personal Privacy

Project Name: **Wolverine World Wide Tannery**

Submittal Date/Time: 12/29/2017 10:00

Collection Date/Time: 12/27/2017 17:28

SDG#: **Ex. 6 - Personal Privacy**

CAT No.	Analysis Name	CAS Number	Result	Method Detection Limit	Dilution Factor
Misc. Organics		EPA 537 Version 1.1	ng/l	ng/l	
14070	NEtFOSAA	2991-50-6	N.D.	2	1
	NEtFOSAA is the acronym for N-ethyl perfluorooctanesulfonamidoacetic Acid.				
14070	NMeFOSAA	2355-31-9	N.D.	2	1
	NMeFOSAA is the acronym for N-methyl perfluorooctanesulfonamidoacetic Acid.				
14070	Perfluorobutanesulfonate	375-73-5	5 J	2	1
14070	Perfluorodecanoic acid	335-76-2	N.D.	2	1
14070	Perfluorododecanoic acid	307-55-1	N.D.	2	1
14070	Perfluoroheptanoic acid	375-85-9	N.D.	2	1
14070	Perfluorohexanesulfonate	355-46-4	3 J	2	1
14070	Perfluorohexanoic acid	307-24-4	N.D.	2	1
14070	Perfluorononanoic acid	375-95-1	N.D.	2	1
14070	Perfluoro-octanesulfonate	1763-23-1	24	2	1
14070	Perfluorooctanoic acid	335-67-1	9	2	1
14070	Perfluorotetradecanoic acid	376-06-7	N.D.	3	1
14070	Perfluorotridecanoic acid	72629-94-8	N.D.	2	1
14070	Perfluoroundecanoic acid	2058-94-8	N.D.	2	1

Sample Comments

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
14070	Full List PFAS - DW	EPA 537 Version 1.1	1	18010003	01/12/2018 22:06	Marissa C Drexinger	1
14381	DW PFAS Prep	EPA 537 Version 1.1	2	18010003	01/10/2018 09:00	Pamela Rothharpt	1

Environmental Analysis Request/Chain of Custody



Lancaster Laboratories
Environmental

For Eurofins Lancaster Laboratories Environmental use only

Acct. # 10459 Group # 1891712 Sample # 9388817-45

COC # 540817

Client Information				Matrix				Analysis Requested												For Lab Use Only																																																																																																							
Client:		Acct. #:		<input type="checkbox"/> Tissue <input type="checkbox"/> Ground <input type="checkbox"/> Surface <input type="checkbox"/> Potable <input checked="" type="checkbox"/> NPDES <input type="checkbox"/> Water <input type="checkbox"/> Other:		Preservation Codes												FSC: _____																																																																																																									
Project Name/#:		PWSID #:				<table border="1"> <tr><td>0</td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td></tr> </table>												0																SCR#: _____																																																																																									
0																																																																																																																											
Project Manager:		P.O. #:		<input type="checkbox"/> Sediment <input type="checkbox"/> Soil <input type="checkbox"/> Composite		<table border="1"> <tr><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td></tr> </table>																												Preservation Codes H=HCl T=Thiosulfate N=HNO ₃ B=NaOH S=H ₂ SO ₄ O=Other																																																																																									
Sampler:		Quote #:																																																																																																																									
State where samples were collected:		For Compliance:		<input type="checkbox"/> Grab <input type="checkbox"/> Composite		<table border="1"> <tr><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td></tr> </table>																												Remarks 																																																																																									
MI		Yes <input type="checkbox"/> No <input checked="" type="checkbox"/>																																																																																																																									
Sample Identification <table border="1"> <thead> <tr> <th>Date</th> <th>Time</th> <th>Grab</th> <th>Composite</th> </tr> </thead> <tbody> <tr><td>12-27-17</td><td>1719</td><td>X</td><td></td></tr> <tr><td>1</td><td>1723</td><td>X</td><td></td></tr> <tr><td>1</td><td>1728</td><td>X</td><td></td></tr> <tr><td>1</td><td>1809</td><td>X</td><td></td></tr> <tr><td>1</td><td>1811</td><td>X</td><td></td></tr> </tbody> </table>				Date	Time	Grab	Composite	12-27-17	1719	X		1	1723	X		1	1728	X		1	1809	X		1	1811	X		Total # of Containers		<table border="1"> <tr><td>1</td><td>X</td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td></tr> <tr><td>2</td><td>X</td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td></tr> <tr><td>2</td><td>X</td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td></tr> <tr><td>1</td><td>X</td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td></tr> <tr><td>2</td><td>X</td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td></tr> </table>												1	X															2	X															2	X															1	X															2	X																
Date	Time	Grab	Composite																																																																																																																								
12-27-17	1719	X																																																																																																																									
1	1723	X																																																																																																																									
1	1728	X																																																																																																																									
1	1809	X																																																																																																																									
1	1811	X																																																																																																																									
1	X																																																																																																																										
2	X																																																																																																																										
2	X																																																																																																																										
1	X																																																																																																																										
2	X																																																																																																																										
Turnaround Time (TAT) Requested (please circle) Standard <input type="checkbox"/> Rush <input checked="" type="checkbox"/> (Rush TAT is subject to laboratory approval and surcharge.) Date results are needed: <u>5-Day</u> E-mail address: <u>Britchie@MannikSmithGroup.com</u>				Relinquished by <u>[Signature]</u>		Date <u>12-27-17</u> Time <u>2:00</u>		Received by _____		Date _____ Time _____																																																																																																																	
				Relinquished by _____		Date _____ Time _____		Received by _____		Date _____ Time _____																																																																																																																	
				Relinquished by _____		Date _____ Time _____		Received by _____		Date _____ Time _____																																																																																																																	
				Relinquished by _____		Date _____ Time _____		Received by _____		Date _____ Time _____																																																																																																																	
				Relinquished by _____		Date _____ Time _____		Received by _____		Date _____ Time _____																																																																																																																	
Data Package Options (circle if required) Type I (EPA Level 3) <input type="checkbox"/> Level IV <input checked="" type="checkbox"/> Type VI (Raw Data Only) Equivalent/non-CLP <input type="checkbox"/> Type III (Reduced non-CLP) <input type="checkbox"/> NJ DKQP <input type="checkbox"/> TX TRRP-13 NYSDEC Category A or B <input type="checkbox"/> MA MCP <input type="checkbox"/> CT RCP <input type="checkbox"/>				EDD Required? Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> If yes, format: _____ Site-Specific QC (MS/MSD/Dup)? Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> (If yes, indicate QC sample and submit triplicate sample volume.)		Relinquished by Commercial Carrier: UPS _____ FedEx X <input checked="" type="checkbox"/> Other _____ Temperature upon receipt <u>1.3</u> °C																																																																																																																					

Ex. 6 - Personal Privacy



Lancaster Laboratories
Environmental

Sample Administration Receipt Documentation Log

Doc Log ID: 205017



Group Number(s): 1891712

Client: MSG/TETRA TECH

Delivery and Receipt Information

Delivery Method:	<u>Fed Ex</u>	Arrival Timestamp:	<u>12/29/2017 10:00</u>
Number of Packages:	<u>4</u>	Number of Projects:	<u>1</u>
State/Province of Origin:	<u>MI</u>		

Arrival Condition Summary

Shipping Container Sealed:	Yes	Sample IDs on COC match Containers:	Yes
Custody Seal Present:	Yes	Sample Date/Times match COC:	Yes
Custody Seal Intact:	Yes	VOA Vial Headspace \geq 6mm:	N/A
Samples Chilled:	Yes	Total Trip Blank Qty:	0
Paperwork Enclosed:	Yes	Air Quality Samples Present:	No
Samples Intact:	Yes		
Missing Samples:	No		
Extra Samples:	No		
Discrepancy in Container Qty on COC:	No		

Unpacked by Wendy Wakeley (1669) at 11:06 on 12/29/2017

Samples Chilled Details

Thermometer Types: DT = Digital (Temp. Bottle) IR = Infrared (Surface Temp) All Temperatures in °C.

Cooler #	Thermometer ID	Corrected Temp	Therm. Type	Ice Type	Ice Present?	Ice Container	Elevated Temp?
1	DT146	3.6	DT	Wet	Y	Bagged	N
2	DT146	1.3	DT	Wet	Y	Bagged	N
3	DT146	0.4	DT	Wet	Y	Bagged	N
4	DT146	1.0	DT	Wet	Y	Bagged	N



Lancaster Laboratories
Environmental

Explanation of Symbols and Abbreviations

The following defines common symbols and abbreviations used in reporting technical data:

BMQL	Below Minimum Quantitation Level	mg	milligram(s)
C	degrees Celsius	mL	milliliter(s)
cfu	colony forming units	MPN	Most Probable Number
CP Units	cobalt-chloroplatinate units	N.D.	non-detect
F	degrees Fahrenheit	ng	nanogram(s)
g	gram(s)	NTU	nephelometric turbidity units
IU	International Units	pg/L	picogram/liter
kg	kilogram(s)	RL	Reporting Limit
L	liter(s)	TNTC	Too Numerous To Count
lb.	pound(s)	µg	microgram(s)
m3	cubic meter(s)	µL	microliter(s)
meq	milliequivalents	umhos/cm	micromhos/cm
<	less than		
>	greater than		
ppm	parts per million - One ppm is equivalent to one milligram per kilogram (mg/kg) or one gram per million grams. For aqueous liquids, ppm is usually taken to be equivalent to milligrams per liter (mg/l), because one liter of water has a weight very close to a kilogram. For gases or vapors, one ppm is equivalent to one microliter per liter of gas.		
ppb	parts per billion		
Dry weight basis	Results printed under this heading have been adjusted for moisture content. This increases the analyte weight concentration to approximate the value present in a similar sample without moisture. All other results are reported on an as-received basis.		

Analytical test results meet all requirements of the associated regulatory program (i.e., NELAC (TNI), DoD, and ISO 17025) unless otherwise noted under the individual analysis.

Measurement uncertainty values, as applicable, are available upon request.

Tests results relate only to the sample tested. Clients should be aware that a critical step in a chemical or microbiological analysis is the collection of the sample. Unless the sample analyzed is truly representative of the bulk of material involved, the test results will be meaningless. If you have questions regarding the proper techniques of collecting samples, please contact us. We cannot be held responsible for sample integrity, however, unless sampling has been performed by a member of our staff.

This report shall not be reproduced except in full, without the written approval of the laboratory.

Times are local to the area of activity. Parameters listed in the 40 CFR Part 136 Table II as "analyze immediately" are not performed within 15 minutes.

WARRANTY AND LIMITS OF LIABILITY - In accepting analytical work, we warrant the accuracy of test results for the sample as submitted. THE FOREGOING EXPRESS WARRANTY IS EXCLUSIVE AND IS GIVEN IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED. WE DISCLAIM ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED, INCLUDING A WARRANTY OF FITNESS FOR PARTICULAR PURPOSE AND WARRANTY OF MERCHANTABILITY. IN NO EVENT SHALL EUROFINS LANCASTER LABORATORIES ENVIRONMENTAL, LLC BE LIABLE FOR INDIRECT, SPECIAL, CONSEQUENTIAL, OR INCIDENTAL DAMAGES INCLUDING, BUT NOT LIMITED TO, DAMAGES FOR LOSS OF PROFIT OR GOODWILL REGARDLESS OF (A) THE NEGLIGENCE (EITHER SOLE OR CONCURRENT) OF EUROFINS LANCASTER LABORATORIES ENVIRONMENTAL AND (B) WHETHER EUROFINS LANCASTER LABORATORIES ENVIRONMENTAL HAS BEEN INFORMED OF THE POSSIBILITY OF SUCH DAMAGES. We accept no legal responsibility for the purposes for which the client uses the test results. No purchase order or other order for work shall be accepted by Eurofins Lancaster Laboratories Environmental which includes any conditions that vary from the Standard Terms and Conditions, and Eurofins Lancaster Laboratories Environmental hereby objects to any conflicting terms contained in any acceptance or order submitted by client.

Data Qualifiers

Qualifier	Definition
C	Result confirmed by reanalysis
D1	Indicates for dual column analyses that the result is reported from column 1
D2	Indicates for dual column analyses that the result is reported from column 2
E	Concentration exceeds the calibration range
J (or G, I, X)	Estimated value \geq the Method Detection Limit (MDL or DL) and $<$ the Limit of Quantitation (LOQ or RL)
P	Concentration difference between the primary and confirmation column $>40\%$. The lower result is reported.
U	Analyte was not detected at the value indicated
V	Concentration difference between the primary and confirmation column $>100\%$. The reporting limit is raised due to this disparity and evident interference.
W	The dissolved oxygen uptake for the unseeded blank is greater than 0.20 mg/L.
Z	Laboratory Defined - see analysis report

Additional Organic and Inorganic CLP qualifiers may be used with Form 1 reports as defined by the CLP methods.

Qualifiers specific to Dioxin/Furans and PCB Congeners are detailed on the individual Analysis Report.